



CERTIFICATE OF ANALYSIS

PRODUCT PO0158A
PLATE COUNT AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1151025

EXPIRY DATE 2011.01.06

| Delivery/Customer information |
|-------------------------------|
| Date Printed 2013.11.11 |
| Delivery No. |
| Customer |
| Customer Order number |

| Physical Characteristics | Results | Specification |
|--|----------|---------------------------|
| Appearance | Conforms | Straw 1 |
| pH (25°C) | 7.1 | 6.8 - 7.2 |
| Fill Volume/Weight | Conforms | Conforms to specification |
| Sterility @ 22°C & 32°C ± 2°C for 5 days | Conforms | Conforms to specification |

| Microbiological Performance | Control c.f.u | Test c.f.u | Specification |
|----------------------------------|---------------|------------|---------------------------|
| Escherichia coli ATCC@25922 | 52 | 55 | Straw colonies |
| Bacillus subtilis ATCC@6633 | 44 | 38 | Irregular, straw colonies |
| Staphylococcus aureus ATCC@6538P | 60 | 55 | Cream colonies |

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd