



G & M Procter Ltd. Certificate of Analysis

PRODUCT EB0209C
PURIFIED WATER
200X5ML IN BIJOU

LOT NUMBER 1178893
EXPIRY DATE 2012.06.29
PRODUCTION DATE 2011.07.04
TEST DATE 2011.06.30
REPORTING DATE 2011.07.10

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Water white	Water white	Appearance and colour
pH (25°C)	9.9	1.0 - 14.0	pH
Fill Volume/Weight	5.1g	5.0 - 5.3g	Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	No growth	No growth	Ster.at 22,32,37 & 44°C

MICROBIOLOGICAL PERFORMANCE

For the target organism both inocula must achieve a colony count of 50-150 cfu. Recovery after holding at 25 degrees C +/- 2 degrees C for 30 minutes shall be between 50% - 150% of the initial count.

Target Organism	Initial cfu	Post holding period cfu	Specification	Accredited Method Reference
Escherichia coli ATCC®8739	103	96	As described above	Maintenance of viability

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Ian Snowball
Ian Snowball
Product Performance Manager
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2013.11.11

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

A handwritten signature in black ink that reads "Ian Snowball".

Ian Snowball
Product Performance Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX