



CERTIFICATE OF ANALYSIS

PRODUCT **MP1571**
DENIM BLUE AGAR (CHROMOGENIC MRSA)

LOT NUMBER 1355907

EXPIRY DATE 25.07.2013
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

08.11.2013

Delivery No.

Customer

106343

ZDELETE Alpha Supplies

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Opaque buff	Conforms to Specification
pH	7.40	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Proteus mirabilis ATCC® 12453	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 33591	Growth; ppt to 1mm blue colonies	Conforms to Specification
Staphylococcus aureus ATCC® 43300	Growth; ppt to 1mm blue colonies	Conforms to Specification
Enterococcus faecalis ATCC® 29212	Inhibition	Conforms to Specification
Staphylococcus epidermidis ATCC® 12228	Inhibition	Conforms to Specification
Pseudomonas aeruginosa ATCC® 27853	Inhibition	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager