

Certificate of Analysis

Product Name: V-Lyte Standard B/Salt Bridge
Product Numbers: K825
Lot Number: KB4D01

Production Date: 2014-04-22
Expiration Date: 2015-04-14

Description	Part Number	Component Lot Number
Standard B	JCK825B	457914
Salt Bridge	JCK825S	457918

Specification	Specification Limit	Result
Standard B		
Appearance	Clear, no precipitates	Complies
Sodium	9.82 - 10.04 mmol/L	9.93
Potassium	0.198 - 0.202 mmol/L	0.199
Chloride	5.86 - 6.00 mmol/L	5.92
pH at 25 °C	9.30 - 9.50	9.43
Ave. Dissolved Oxygen Conc. In-Process	<3.5 PPM	2.4
Ave. Dissolved Oxygen Conc. End Use	<6.5 PPM	5.0
Air Volume (cc)	<3.5	0.7

Salt Bridge		
Appearance	Clear, no precipitates	Complies
Sodium	<0.100 mmol/L	0.005
Potassium	118.0 - 122.0 mmol/L	119.1
Chloride	118.33 - 122.33 mmol/L	120.16
pH at 25 °C	3.5 - 3.9	3.7
Ave. Dissolved Oxygen Conc. In-Process	<3.5 PPM	1.9
Ave. Dissolved Oxygen Conc. End Use	<6.5 PPM	3.8
Air Volume (cc)	<3.5	0.4

For in-vitro diagnostic use.



Quality Representative

Date: 2014-05-05