

CERTIFICATE OF ANALYSIS

PRODUCT PP2452
PLATE COUNT AGAR (APHA)

LOT NUMBER 1295258

EXPIRY DATE 2014.05.23

RELEASE DATE 2014.03.30

Physical Testing	Results	Specification
------------------	---------	---------------

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Colourless
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	7.0	6.8 - 7.2
Sterility @ 30°C for 3 days /sample plan	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiology Testing	Result	Specification	Method
----------------------	--------	---------------	--------

Escherichia coli ATCC® 11775	Pass	Cream colonies	QC27
------------------------------	------	----------------	------

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Paul Costello

Paul Costello

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.



Accreditation
Number 3186



Accreditation No. 3186

MANUFACTURED AND TESTED BY:

Thermo Fisher Scientific, Microbiology Division

20 Dalgleish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au