



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT TV5014E
SIM MEDIUM

LOT NUMBER 1322555

EXPIRY DATE 2013.10.21

TYPICAL FORMULATION

	grams per litre
Tryptone	20.00000
Peptone	6.10000
Ferrous ammonium sulphate	0.20000
Sodium thiosulphate	0.20000
Agar	3.50000

General Characteristics	Results	Specification
Colour	Conforms	Sun yellow
Appearance	Conforms	Transparent
pH	7.3	7.1 - 7.5
Packaging/Presentation	Conforms	Conforms to specification
Filling Format	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	No Contamination

Microbiological Performance

Control strains	Control	Test	Growth/Motility/Indole/H2S
Escherichia coli ATCC®25922	1 Col	Conforms	+ / + / + / -
Salmonella typhimurium ATCC®14028	1 Col	Conforms	+ / + / - / +
Shigella flexneri ATCC®12022	1 Col	Conforms	+ / - / - / -

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature