



Certificate of Quality

Product Name: Charcoal Bld Agar (Regan Lowe) 10/PK
Lot Number: 354260

Product Number: R01298
Expiration Date: 2013-07-24
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bordetella pertussis	12742	Good Growth
Klebsiella pneumoniae	27736	No Growth
Staphylococcus aureus	25923	No Growth
Streptococcus pneumoniae	6305	No Growth

Physical Characteristics:

Appearance: Black to Black-Red, Opaque Agar
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor