

Certificate of Quality

Product Name: CLED Agar 10/PK
Lot Number: 723627

Product Number: R01312
Expiration Date: 2009-01-29
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Proteus vulgaris	8427	Swarming Blue Colonies, Inhibited
Escherichia coli	25922	Yellow Opaque Colonies
Staphylococcus aureus	25923	Deep Yellow Colonies

Physical Characteristics:

Appearance: Blue-Green, Clear Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor