

Certificate of Quality

Product Name: Blood Agar w/Amp(10mcg) 10/PK
Lot Number: 754553

Product Number: R01226
Expiration Date: 2009-05-13
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Aeromonas hydrophila	7965	Good Growth, Beta Hemolysis
Escherichia coli	25922	No Growth
Proteus mirabilis	12453	No Growth

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor