

# Certificate of Quality

**Product Name:** Candida BCG Agar 10/PK  
**Lot Number:** 756428

**Product Number:** R01282  
**Expiration Date:** 2009-06-02  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Candida albicans	10231	Good Growth, Yellow Zone
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth

**Physical Characteristics:**

Appearance: Blue-Green, Clear Agar  
pH @ 20-25°C: 6.1 +/- 0.2

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor