

Certificate of Quality

Product Name: Charcoal Selective Medium 10/PK
Lot Number: 801917

Product Number: R01294
Expiration Date: 2009-10-08
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Campylobacter jejuni	33291	Good Growth
Campylobacter coli	33559	Good Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth
Pseudomonas aeruginosa	27853	No Growth
Proteus mirabilis	12453	No Growth
Cryptococcus neoformans	34877	No Growth

Physical Characteristics:

Appearance: Black, Opaque Agar
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor