

**CERTIFICATE OF ANALYSIS**

PRODUCT BM0104T
BUFFERED PEPTONE WATER READY BAG 4X3 L

LOT NUMBER 2477209

EXPIRY DATE 29.02.2020

DATE OF MANUFACTURE 15.02.2019

Delivery/Customer information

Date Printed
 2021.07.27
 Delivery No.

Customer
 Customer Order Number

Physical Characteristics	Results	Specification
Colour	Straw 1-2	Straw 1-2
pH (25°C)	7.3	7.0 - 7.4
Clarity	Clear	Clear
Packaging and Presentation	Conforms	Conforms to specification
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 20-25°C & 30-35°C for 5 days	Satisfactory	Satisfactory

Microbiological Performance	Initial Inoculum	Test Result
-----------------------------	------------------	-------------

Tested as a non-selective pre-enrichment broth**Aerobic incubation at 37 ± 2°C for 18 ± 2 hours**

<i>Salmonella nottingham</i> NCTC7832	44	Turbid growth
<i>Escherichia coli</i> ATCC®11775	54	Turbid growth

Testing has been performed in accordance with ISO11133:2014**Aerobic incubation at 37 ± 2°C for 18 ± 2 hours****A satisfactory result is represented by visible growth from an inoculum of 10-100 cfu.**

<i>S. typhimurium</i> ATCC®14028 WDCM00031	36	Turbid growth
<i>S. enteritidis</i> ATCC®13076 WDCM00030	72	Turbid growth
<i>E. coli</i> ATCC®25922 WDCM00013	72	Turbid growth
<i>E. coli</i> ATCC®8739 WDCM00012	43	Turbid growth

Testing has been performed in accordance with ISO22964:2017**Aerobic incubation at 36 ± 2°C for 18 ± 2 hours**

<i>C. sakazakii</i> ATCC®29544 WDCM00214	43	Turbid growth
<i>C. muytjensii</i> ATCC®51329 WDCM00213	34	Turbid growth

Control Medium: Tryptone Soya Agar

Refer to product specification for full details.

A satisfactory sterility test result is represented by no visible microbial growth after incubation for 5 days at 20-25°C and at 30-35°C.



Tested by the Quality Control Laboratory
 OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
 oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT **BM0104T**
LOT NUMBER **BUFFERED PEPTONE WATER READY BAG 4X3 L**
 2477209

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2019.02.26

.....
Shannon Bearpark
Site Quality Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com