

G & M Procter Ltd. Certificate of Analysis

PRODUCT **BO0100M**
PURIFIED WATER (FOR MICROBIOLOGY)
10X100ML IN 100ML VIAL

LOT NUMBER 2678406
EXPIRY DATE 2020.03.13
PACKING DATE 2019.03.13
TEST DATE 2019.03.13
REPORTING DATE 2019.03.19

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Colourless	Colourless	SOP 183 Appearance and colour
pH (25°C)	9.8	1.0 - 14.0	SOP 53 pH
Fill Volume/Weight	100.3g	100.0 - 101.5g	SOP 154 Fill volume weight check
Minimal fluid filtration	No growth	No growth	SOP 311 Contamination Check at 32°C on Diluents

MICROBIOLOGICAL PERFORMANCE

For the target organism both inocula must achieve a colony count of 50-150 cfu. Recovery after holding at 25 degrees C +/- 2 degrees C for 30 minutes shall be between 50% - 150% of the initial count.

Target Organism	Initial cfu	Post holding period cfu	Specification	Accredited Method Reference
Escherichia coli ATCC®8739	122	119	As described above	SOP 139 Maintenance of viability

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Ian Snowball
Ian Snowball
Quality Manager
G & M Procter Ltd

The test laboratory of Thermo Fisher Scientific is accredited by UK accreditation authority UKAS as registered test laboratory 2727 according to EN ISO/IEC 17025 for the performance testing of media for microbiology.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road, Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2019.03.19

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.



Ian Snowball
Quality Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX