

**CERTIFICATE OF ANALYSIS**

**PRODUCT** CM0485R  
**VIOLET RED BILE GLUCOSE AGAR 2.5kg**

**LOT NUMBER** 1416752

**EXPIRY DATE** 2018.11.30

**DATE OF MANUFACTURE** 2013.10.22

## Delivery/Customer information

Date Printed  
 2018.07.05  
 Delivery No.  
 Customer  
 Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw/pink powder	Straw/pink powder
Colour on reconstitution	Purple	Purple
pH (25°C)	7.5	7.2 - 7.6
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery %	Test Description
<b>Aerobic incubation at 37°C for 24 hours</b>				
<b>Pour plate technique</b>				
<i>Klebsiella pneumoniae</i> ATCC®29665	82	70	85	Purple/pink cols & halo
<i>Proteus mirabilis</i> ATCC®12453	143	120	84	Purple/pink cols & halo
<b>Surface plate technique</b>				
<i>Shigella sonnei</i> ATCC®25931	67	50	75	Irregular, purple/pink cols & halo
<i>Enterobacter aerogenes</i> ATCC®13048	41	39	95	Purple/pink, mucoid cols & halo
<i>Pseudomonas aeruginosa</i> ATCC®9027	31	25	81	Straw cols, no halo
<i>Staphylococcus aureus</i> ATCC®6538	>1E+04	0	-	No growth
<b>Aerobic incubation at 37°C for 24 hours</b>				
<b>Pour plate technique</b>				
<i>E.coli</i> ATCC®25922 WDCM00013	91	70	77	Purple/pink cols & halo
<i>E.coli</i> ATCC®8739 WDCM00012	71	75	106	Purple/pink cols & halo
<i>S.typhimurium</i> ATCC®14028 WDCM00031	42	39	93	Purple/pink cols & halo
<b>Surface plate technique</b>				
<i>E.faecalis</i> ATCC®29212 WDCM00087	>1E+04	0	-	No growth
<i>E.faecalis</i> ATCC®19433 WDCM00009	>1E+04	0	-	No growth

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 50% of the control medium.

Refer to product specification for full details.



Tested by the Product Performance Laboratory  
 OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England  
 oxid@thermofisher.com www.oxid.com




## CERTIFICATE OF ANALYSIS

**PRODUCT** CM0485R  
**VIOLET RED BILE GLUCOSE AGAR 2.5kg**  
**LOT NUMBER** 1416752

---

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.  
Lot Accepted. 2013.10.28

  
.....  
Carol MacBride  
Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.  
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Product Performance Laboratory  
**OXOID LIMITED**  
Wade Road, Basingstoke, Hampshire RG24 8PW. England  
[oxoid@thermofisher.com](mailto:oxoid@thermofisher.com) [www.oxoid.com](http://www.oxoid.com)