



Quality Assurance Certificate

Code: CM1050
Product: CHROMOGENIC UTI MEDIUM (CLEAR)
Lot: 2896842
Expiry: 2022.12.31
Components:

PHYSICAL CHARACTERISTICS: As appropriate tests on moisture, appearance, pH, clarity and gel strength were satisfactory.

BIOLOGICAL PERFORMANCE: Satisfactory performance of:

Escherichia coli	ATCC 8739
Enterobacter aerogenes	ATCC 13048
Proteus mirabilis	NCTC 10975
Enterococcus faecalis	ATCC 29212
Staphylococcus aureus	ATCC 25923
Citrobacter freundii	NCTC 8581

Date: 23 November 2019

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.

Carissa Courtney

.....
Carissa Courtney
Director Quality, EMEA

