



# CERTIFICATE OF ANALYSIS

**PRODUCT** DD0005B  
**X+V FACTOR 5x 50 DISCS**

**LOT NUMBER** 2341645

**EXPIRY DATE** 2019.06.30

**DATE OF MANUFACTURE** 2018.05.26

Delivery/Customer information	
Date Printed	2018.06.12
Delivery No.	
Customer	
Customer Order Number	

## Microbiological Performance

### Diagnostic characteristics

	Zone Size (mm)	Limits (mm)
<i>Haemophilus influenzae</i> ATCC®19418	19	14 - 25
<i>Haemophilus influenzae</i> ATCC®9334	20	14 - 25
<i>Haemophilus influenzae</i> ATCC®49247	23	14 - 25
<i>Haemophilus parainfluenzae</i> ATCC®33392	48	35 - 49

Quality control ranges are internally derived based on historical data.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2018.06.11

Shannon Bearpark  
Site Quality Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



**Tested by the Quality Control Laboratory**  
**OXOID LIMITED**  
 Wade Road, Basingstoke, Hampshire RG24 8PW. England  
 oxoid@thermofisher.com www.oxid.com