

G & M Procter Ltd. Certificate of Analysis

PRODUCT **EB1355B**
0.85% SALINE
100 X 3ML IN MFA TUBE

LOT NUMBER 4316560
EXPIRY DATE 2022.06.03
PACKING DATE 2021.06.03
TEST DATE 2021.06.03
REPORTING DATE 2021.06.09

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Colourless	Colourless	SOP 183 Appearance and colour
pH (25°C)	7.1	1.0 - 14.0	SOP 53 pH
Fill Volume/Weight	3.3g	3.0 - 3.3g	SOP 154 Fill volume weight check
Minimal fluid filtration	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

Target Organism	Initial cfu	Post holding period cfu	Specification	Accredited Method Reference
Escherichia coli ATCC®8739	91	77	As described above	SOP 139 Maintenance of viability

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball
 Ian Snowball
 Quality Manager
 G & M Procter Ltd

The test laboratory of Thermo Fisher Scientific is accredited by UK accreditation authority UKAS as registered test laboratory 2727 according to EN ISO/IEC 17025 for the performance testing of media for microbiology.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
 Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727