

**CERTIFICATE OF ANALYSIS**

PRODUCT ID0580M
OBIS - PYR 60 TEST KITS

LOT NUMBER 2223900

EXPIRY DATE 2019.10.31

DATE OF MANUFACTURE 2017.10.03

Delivery/Customer information

Date Printed
2018.06.25

Delivery No.

Customer
Customer Order Number

Physical Characteristics	Results	Specification
Appearance Buffer Reagent ID0100M	Colourless with no odour	Colourless with no odour
Appearance Colour Developer ID0200M	Pale yellow with no odour	Pale yellow with no odour
Microbiological Performance	Initial Inoculum	Test Result
<i>Citrobacter freundii</i> ATCC®8090	>1E+04	Positive reaction at 20 seconds
<i>Enterococcus faecalis</i> ATCC®29212	>1E+04	Positive reaction at 20 seconds
<i>Streptococcus pyogenes</i> ATCC®19615	>1E+04	Positive reaction at 20 seconds
<i>Streptococcus agalactiae</i> ATCC®13813	>1E+04	Negative reaction at 20 seconds
<i>Salmonella typhimurium</i> ATCC®14028	>1E+04	Negative reaction at 20 seconds
<i>Salmonella enteritidis</i> ATCC®13076	>1E+04	Negative reaction at 20 seconds

Refer to product specification for full details.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2017.10.05

Shannon Bearpark
Site Quality Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096