

Certificate of Analysis

Product Name: LDH-L, 20X20ML
Product Numbers: TR20015
Lot Number: 317674

Date of Manufacture: 05-FEB-13
Expiration Date: 2015-04-30
Storage: 2 - 8°C

Specification	Specification Limit	Result
Appearance (Powder)	White to off-white free-flowing powder	Complies
Appearance (Reconstituted)	Clear, colorless solution essentially free of particulate matter	Complies
Data-Trol Normal	Lower Limit	99 U/L
Data-Trol Normal	Upper Limit	159 U/L
Data-Trol Normal	Result	136 U/L
Data-Trol Normal	Lot Number	V36526
Data-Trol Abnormal	Lower Limit	356 U/L
Data-Trol Abnormal	Upper Limit	481 U/L
Data-Trol Abnormal	Result	437 U/L
Data-Trol Abnormal	Lot Number	V36528
Linearity	Linear to 1000 U/L	Complies

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications, and has been released.

Connie Carroll

Quality Representative

Date: 2013-04-17