

Daliyary/Customar information

CERTIFICATE OF ANALYSIS

DATE OF MANUFACTURE	2018.06.19	
EXPIRY DATE	2019.07.31	Customer Order Number
LOT NUMBER	2356763	Customer
PRODUCT	MA0102F VANCOMYCIN 256 M.I.C.E. (50 Pack)	Date Printed 2018.06.26 Delivery No.
		Derivery/Customer information

Microbiological Performance

Mueller-Hinton Agar

Incubated at 35-37°C in air for 16-18 hours		
	MIC	Limits
	(µg/mL)	(µg/mL)
Staphylococcus aureus ATCC®29213	1.000	0.375 - 2.000
Enterococcus faecalis ATCC®29212	3.00	0.75 - 4.00
Mueller-Hinton Agar enriched with 5% v/v sh Incubated at 35-37°C in 5% CO2 for 20-24 ho Streptococcus pneumoniae ATCC®49619	-	0.090 - 0.500
Sirepiococcus pricumonitie AICC@+)01)	0.120	0.070 - 0.500

4 < 40.

The result is expressed as the mean result rounded to the nearest half or full MIC value. Where the mean MIC value falls between two MIC values the highest value is reported.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release. Lot Accepted.

Branpark

Shannon Bearpark Site Quality Manager

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory OXOID LIMITED Wade Road, Basingstoke, Hampshire RG24 8PW. England oxoid@thermofisher.com www.oxoid.com