



CERTIFICATE OF ANALYSIS

PRODUCT **MBB0004**
PEPTONE WATER 0.1% 99 ML PCK 50
125ML PP BOTTLES CR00387

LOT NUMBER 3293

EXPIRY DATE 19.10.2004
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

21.06.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear liquid	Conforms to Specification
pH	6.20	6.10 - 6.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance	Expected Results	Specification
Organism Tested		
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Salmonella typhimurium ATCC® 14028	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor