



CERTIFICATE OF ANALYSIS

PRODUCT **MBB1802**
RAKA-RAY AGAR (MODIFIED) 250 ML CUSTOM
500 ML CLEAR GLASS CR00446

LOT NUMBER 7177

EXPIRY DATE 16.10.2007
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

27.06.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Straw coloured	Conforms to Specification
pH	5.20	5.20 - 5.60
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Lactobacillus fermentans ATCC® 9338	Good growth; white cream colonies	Conforms to Specification
Saccharomyces cerevisiae ATCC® 9763	Inhibition	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor