



CERTIFICATE OF ANALYSIS

PRODUCT **MP0232**
BHI AGAR W/CHLORAMPHENICOL,GENTAMICIN

LOT NUMBER 2413801

EXPIRY DATE 21.12.2018
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

03.10.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear amber	Conforms to Specification
pH	7.30	7.20 - 7.60
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Trichophyton mentagrophytes ATCC® 9533	Growth	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Candida albicans ATCC® 10231	Growth	Conforms to Specification
A. brasiliensis frm. niger ATCC® 16404	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor