



CERTIFICATE OF ANALYSIS

PRODUCT **MP0299**
CAMPY-CEFEX AGAR 10/PK
100MM PACK OF 10

LOT NUMBER 3326070

EXPIRY DATE 03.09.2021
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

13.07.2021

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Brown to brown red	Conforms to Specification
pH	7.10	6.80 - 7.20
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance

Organism Tested	Expected Results	Specification
Campylobacter jejuni ATCC® 33291	Good growth	Conforms to Specification
Pseudomonas aeruginosa ATCC® 27853	No growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Peter Kornherr
Quality Assurance Manager