



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP0307**  
**CAMPYLOBACTER AGAR (CVA)**

**LOT NUMBER**            6076

**EXPIRY DATE**            12.05.2006  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.07.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.20	7.20 - 7.60
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition (partial)	Conforms to Specification
Proteus mirabilis ATCC® 43071	Inhibition	Conforms to Specification
Campylobacter jejuni ATCC® 33291	Growth	Conforms to Specification
Serratia marcescens ATCC® 8100	Inhibition (partial)	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

*Sandra DeMille*

Sandra DeMille  
Quality Control Supervisor