



CERTIFICATE OF ANALYSIS

PRODUCT **MP0330**
 CHOCOLATE ENRICHED

LOT NUMBER 3108479

EXPIRY DATE 08.10.2020
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

20.07.2020

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Chocolate brown	Conforms to Specification
pH	7.20	7.00 - 7.40
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance

Organism Tested	Expected Results	Specification
Neisseria meningitidis ATCC® 13090	Growth	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43069	Growth	Conforms to Specification
Haemophilus parainfluenzae ATCC® 7901	Growth	Conforms to Specification
Haemophilus influenzae ATCC® 49766	Growth	Conforms to Specification
Haemophilus influenzae ATCC® 10211	Growth	Conforms to Specification
Neisseria gonorrhoeae ATCC® 43070	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor