

CERTIFICATE OF ANALYSIS

PRODUCT

MP0351

COLUMBIA BLOOD AGARW/5% SHEEP BLOOD

LOT NUMBER

2341083

EXPIRY DATE

19.07.2018

(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

18.05.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.30	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ $35 ^{\circ}\text{C} \pm 2 ^{\circ}\text{C}$ for 3days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922 Staphylococcus aureus ATCC® 25923 Streptococcus pyogenes ATCC® 19615 Streptococcus pneumoniae ATCC® 6305	Growth Growth Growth, beta hemolysis Growth, alpha hemolysis	Conforms to Specification Conforms to Specification Conforms to Specification Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille

Quality Control Supervisor