



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP1244**  
**BRILLIANCE GBS AGAR**  
**100 MM, PACK OF 10**

**LOT NUMBER**            2374586

**EXPIRY DATE**            11.09.2018  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

20.07.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear	Conforms to Specification
pH	6.80	6.60 - 7.00
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Streptococcus agalactiae ATCC® 12386	pink colonies	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

*Sandra DeMille*

Sandra DeMille  
Quality Control Supervisor