



CERTIFICATE OF ANALYSIS

PRODUCT **MP1312**
MACCONKEY AGAR #3 W/ CRYSTAL VIOLET

LOT NUMBER 3171117

EXPIRY DATE 18.01.2021
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

29.10.2020

Delivery No.

Customer

Customer Order number

| Physical Characteristics | Results | Specification |
|------------------------------------|------------|---------------------------|
| Appearance | Clear pink | Conforms to Specification |
| pH | 7.00 | 6.90 - 7.30 |
| Packaging and Presentation | Conforms | Conforms to specification |
| Sterility @ 35 °C ± 2°C for 3 days | Conforms | Conforms to specification |

Microbiological Performance

| Organism Tested | Expected Results | Specification |
|------------------------------------|------------------------------------------|---------------------------|
| Escherichia coli ATCC® 25922 | Growth, pink colonies | Conforms to Specification |
| Proteus mirabilis ATCC® 12453 | Growth colorless cols swarming inhibited | Conforms to Specification |
| Salmonella typhimurium ATCC® 14028 | Growth, colourless colonies. | Conforms to Specification |
| Enterococcus faecalis ATCC® 29212 | Inhibition | Conforms to Specification |
| Staphylococcus aureus ATCC® 25923 | Inhibition | Conforms to Specification |

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Peter Kornherr
Quality Assurance Manager