



CERTIFICATE OF ANALYSIS

PRODUCT **MP1369**
MUELLER HINTON AGAR PLUS
100MM

LOT NUMBER 2389350

EXPIRY DATE 06.11.2018
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

17.08.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear amber	Conforms to Specification
pH	7.30	7.20 - 7.40
Depth of media	4.4	3.5 - 5.0
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Escherichia coli ATCC® 35218	Growth	Conforms to Specification
Pseudomonas aeruginosa ATCC® 27853	Growth	Conforms to Specification
Enterococcus faecalis ATCC® 33186	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 43300	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor