



CERTIFICATE OF ANALYSIS

PRODUCT **MP1620**
PHENYLETHYL ALCOHOL AGAR

LOT NUMBER 6048

EXPIRY DATE 14.04.2006
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

25.06.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.20	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Proteus mirabilis ATCC® 12453	Inhibition (partial to complete)	Conforms to Specification
Streptococcus pyogenes ATCC® 19615	Growth	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition (partial to complete)	Conforms to Specification
Enterococcus faecalis ATCC® 29212	Growth	Conforms to Specification
Streptococcus pneumoniae ATCC® 6305	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor