

CERTIFICATE OF ANALYSIS

PRODUCT

MT0230

BILE AESCULIN AGAR SLANT 5 ML 16x100MM

LOT NUMBER

2424122

EXPIRY DATE

20.04.2019

(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

25.10.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification	
Appearance pH	Clear dark amber 7.10	Conforms to Specification 6.90 - 7.30	
Sterility @ $35 ^{\circ}\text{C} \pm 2 ^{\circ}\text{C}$ for 3days	Conforms	Conforms to specification	

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922 Enterococcus faecalis ATCC® 29212	Growth Growth, darkening of media	Conforms to Specification Conforms to Specification
Streptococcus pyogenes ATCC® 19615	Inhibition (partial to complete)	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille

Quality Control Supervisor