



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT0513  
VRE SELECTIVE BROTH(BEA) 10 ML 16X125MM  
16 X 125MM

**LOT NUMBER** 2125

**EXPIRY DATE** 03.11.2012  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

21.06.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear amber	Conforms to Specification
pH	7.10	6.90 - 7.30
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Pseudomonas aeruginosa ATCC® 27853	Inhibition	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Enterococcus faecalis ATCC® 29212	Inhibition	Conforms to Specification
Enterococcus faecalis ATCC® 51299	Growth (blackening of broth)	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

*Sandra DeMille*

Sandra DeMille  
Quality Control Supervisor