



CERTIFICATE OF ANALYSIS

PRODUCT MT1567
O-F MEDIUM W/ FRUCTOSE 3ML 13x100MM

LOT NUMBER 2842301

EXPIRY DATE 11.02.2020
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

13.11.2019

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Green stab	Conforms to Specification
pH	6.70	6.60 - 7.00
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Klebsiella pneumoniae ATCC® 13883	Fermentation	Conforms to Specification
Acinetobacter Iwoffii ATCC® 17925	No change	Conforms to Specification
Stenotrophomonas maltophilia ATCC® 13637	Oxidation	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor