



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT1910  
SAB DEX AGAR SLANT W/CHLOR 10ML  
20X125MM

**LOT NUMBER** 2283498

**EXPIRY DATE** 26.07.2018  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.07.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Amber slant	Conforms to Specification
pH	5.80	5.40 - 5.80
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Candida albicans ATCC® 10231	Growth	Conforms to Specification
Trichophyton mentagrophytes ATCC® 9533	Growth	Conforms to Specification
A. brasiliensis frm. niger ATCC® 16404	Growth	Conforms to Specification
Trichophyton rubrum ATCC® 28188	Growth	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition (partial to complete)	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

*Sandra DeMille*

Sandra DeMille  
Quality Control Supervisor