



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT1943  
SELENITE BROTH 10 ML 16x125MM PCK 100

**LOT NUMBER** 2309114

**EXPIRY DATE** 10.11.2018  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

06.07.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	clear yellow broth , oran	Conforms to Specification
pH	7.00	6.90 - 7.30
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Inhibition (partial to complete)	Conforms to Specification
Salmonella typhimurium ATCC® 14028	Growth	Conforms to Specification
Shigella sonnei ATCC® 9290	Variable growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

*Sandra DeMille*

Sandra DeMille  
Quality Control Supervisor