



CERTIFICATE OF ANALYSIS

PRODUCT MT1959
STAPH BROTH W/ OXACILLIN 10ML 16X125mm
16 X 125 MM

LOT NUMBER 8267

EXPIRY DATE 16.12.2008
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

04.07.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear straw colored	Conforms to Specification
pH	7.10	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Staphylococcus aureus ATCC® 29213	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 43300	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor