



CERTIFICATE OF ANALYSIS

PRODUCT MT1966
STARCH GELATIN DEXTROSE 3ML 13x100MM

LOT NUMBER 2017

EXPIRY DATE 18.07.2012
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

05.07.2018

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Pink agar stab	Conforms to Specification
pH	7.40	7.30 - 7.70
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Neisseria gonorrhoeae ATCC® 43069	Positive, medium turns yellow	Conforms to Specification
Neisseria meningitidis ATCC® 13090	Positive, medium turns yellow	Conforms to Specification
Neisseria sicca ATCC® 9913	Positive, medium turns yellow	Conforms to Specification
Neisseria lactamica ATCC® 23970	Positive, medium turns yellow	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Sandra DeMille

Sandra DeMille
Quality Control Supervisor