

CERTIFICATE OF ANALYSIS

PRODUCT MV1035
GRAM DILUTE CARBOL FUCHSIN 4L

LOT NUMBER 1783549

EXPIRY DATE 2019.06.29

RELEASE DATE 2017.07.04

Physical Testing **Results** **Specification**

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Pink
Clarity	Pass	Opaque
Sterility @ 30°C for 2 days	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiological Testing **Result** **Specification** **Method**

Escherichia coli ATCC® 25922	Pass	Gram -ve bacilli (Pink)	QC06
Staphylococcus aureus ATCC® 25923	Pass	Gram +ve cocci (Purple)	QC06

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Paul Costello

Paul Costello

.....
Signatory

MANUFACTURED AND TESTED BY:
Thermo Fisher Scientific, Microbiology Division
20 Dalglish Street, Thebarton, South Australia 5031
Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054
Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au