

CERTIFICATE OF ANALYSIS

PRODUCT PB5067A G.VAGINALIS AGAR + HUMAN BLOOD
LOT NUMBER 2340913
EXPIRY DATE 2018.07.11

General Characteristics	Results	Specification
Colour	Conforms	Signal red
pH	7.1	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Gardnerella vaginalis ATCC®14018	94	94	Out of specification
Proteus mirabilis ATCC®29906	>100,000	Conforms	No growth
Escherichia coli ATCC®25922	1E+04 - 1E+05	0	Total inhibition
Candida albicans ATCC®10231	1E+04 - 1E+05	0	Total inhibition

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.