

CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT **PB5175A**
PASTEURELLA SELECTIVE AGAR

LOT NUMBER 2396574

EXPIRY DATE 2018.10.11

TYPICAL FORMULATION

	grams per litre
Meat Extract, 'Lab Lemco'	10.00000
Peptone	10.00000
Sodium chloride	5.00000
Agar	15.00000
Defibrinated sheep blood	50.00000 ML
Bacitracin	3.50000 MG
Bacitracin	2.00000 MG

General Characteristics	Results	Specification
Colour	Conforms	Traffic red
Appearance	Conforms	Opaque
pH	7.5	7.2 - 7.6
Packaging / Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Pasteurella multocida DSM5281	78	62	1 - 3mm grey shiny, smooth colonies
Streptococcus oralis ATCC®35037	1.0E+04	0	Partial inhibition (<= 100 cfu)

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature