

CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PB5303A MUELLER HINTON AGAR WITH HORSE BLOOD
LOT NUMBER 2338668
EXPIRY DATE 2018.06.08

General Characteristics	Results	Specification
Colour	Conforms	Signal red
Appearance	Conforms	Opaque
pH	7.4	7.2 -7.4
Packaging / Presentation	Conforms	Label & Print check
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Susceptibility Testing	Results(mm)	Limits (mm)
Strep. pneumoniae ATCC®49619 - E15	30	26 -32
Strep. pneumoniae ATCC®49619 - CPD10	34	29 -35
Strep. pneumoniae ATCC®49619 - MXF5	30	24 -30
Strep. pneumoniae ATCC®49619 - ETP10	34	28 -34
Haemophilus influenzae ATCC®49766 - E15	12	10 -16
Haemophilus influenzae ATCC®49766 -CPD10	30	30 -36
Haemophilus influenzae ATCC®49766 - MXF5	32	30 -36

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkkS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133:2014 and registered under D-PL-20190-01-00.