

G & M Procter Ltd. Certificate of Analysis

PRODUCT	PO1210A BRILLIANCE MRSA 2 AGAR 1 PACK OF 10 PLATES
LOT NUMBER	^y 2518104
EXPIRY DATE	2019.01.11
PACKING DATE	2018.11.16
TEST DATE	2018.11.16
REPORTING DATE	2018.11.20

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Pale buff	Pale buff	SOP 178 Appearance and colour
pH (25°C)	7.4	7.1 - 7.5	SOP 53 pH
Fill Weight 19 ± 2 Grams	Conforms	Conforms	SOP 74 Fill volume weight check
Contamination @ 32°C ± 2°C for >72 hours	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification.

For partially inhibited organisms, the test medium must achieve the level of inhibition stated in the specification from the stated inoculum .

For inhibited organisms, the test medium must show no growth from the stated inoculum.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Staphylococcus aureus MRSA ATCC®33591	76	60	Blue cols	Blue cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Ian Snowball
Ian Snowball
Quality Manager
G & M Procter Ltd

The test laboratory of Thermo Fisher Scientific is accredited by UK accreditation authority UKAS as registered test laboratory 2727 according to EN ISO/IEC 17025 for the performance testing of media for microbiology.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

G & M Procter Ltd. Certificate of Analysis

Inhibited Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Pseudomonas aeruginosa ATCC®27853	10000	No growth	No growth	SOP 155 Inhibition
Staph epidermidis ATCC®12228	10000	No growth	No growth	SOP 155 Inhibition
Staph.aureus var.oxford ATCC®9144	10000	No growth	No growth	SOP 155 Inhibition
Proteus mirabilis ATCC®29906	10000	No growth	No growth	SOP 155 Inhibition
Partially Inhibited Organism(s)	Control (cfu)	Test	Specification	Accredited Method Reference
Bacillus licheniformis ATCC®14580	10,000 - 100,000	Pink cols restricted growth	Pink cols restricted growth	SOP 150 Partial Inhibition

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Ian Snowball
Ian Snowball
Quality Manager
G & M Procter Ltd

The test laboratory of Thermo Fisher Scientific is accredited by UK accreditation authority UKAS as registered test laboratory 2727 according to EN ISO/IEC 17025 for the performance testing of media for microbiology.

CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2018.11.20

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.



Ian Snowball
Quality Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX