

CERTIFICATE OF ANALYSIS

PRODUCT PO5001A SABOURAUD GLUCOSE AGAR
LOT NUMBER 2474435
EXPIRY DATE 2019.07.30

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	5.5	5.4 - 5.8
Packaging / Presentation	Conforms	Conforms to specification
Sterility @ 22 & 32 ± 1°C for 5 days	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Candida albicans ATCC®10231 WDCM 00054	76	61	2-3 mm, cream colonies
Aspergillus brasiliensis ATCC®16404 WDCM 00053	43	43	10-20 mm, white mycelium, black spores
Saccharomyces cerevisiae ATCC®9763 WDCM 00058	80	70	2-3 mm, cream colonies

Biological performance tested in accordance with current version of ISO 11133.

Tested in accordance with BP/ EP/ JP/ USP. Clearly visible growth within 5 days for fungi.

For positive strains, colony count is greater than or equal to 70% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.