

CERTIFICATE OF ANALYSIS

PRODUCT PO5096A SAB GLUC AGAR +GENTAM+CHLORAMPH
LOT NUMBER 2346104
EXPIRY DATE 2018.11.23

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	5.7	5.4 - 5.8
Packaging / Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Candida albicans ATCC®10231	55	54	2-3mm, white colonies
Escherichia coli ATCC®25922	1.2E+04	0	Complete inhibition (<= 10 cfu)
Aspergillus brasiliensis ATCC®16404	1 Col	Conforms	Good growth, white mycelium, black spores

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.