

CERTIFICATE OF ANALYSIS

PRODUCT PO5310A BRILLIANCE (TM) MRSA 2 AGAR
LOT NUMBER 2458603
EXPIRY DATE 2019.02.28

General Characteristics	Results	Specification
Colour	Conforms	Light ivory
Appearance	Conforms	Opaque
pH	7.2	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC@33591	66	61	1-2 mm, blue colonies
Staphylococcus aureus ATCC@29213	1E+04 - 1E+05	0	Total inhibition
Pseudomonas aeruginosa ATCC@27853	1E+04 - 1E+05	0	Total inhibition
Bacillus licheniformis ATCC@14580	<10000	Conforms	Growth of small, rose colonies
Proteus mirabilis ATCC@29906	>100,000	Conforms	No growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133:2014 and registered under D-PL-20190-01-00.