

CERTIFICATE OF ANALYSIS

PRODUCT PO5331A **GLUCOSE TRYPTONE AGAR**
LOT NUMBER 3228628
EXPIRY DATE 2021.03.26

General Characteristics	Results	Specification
Colour	Conforms	Blue lilac
Appearance	Conforms	Transparent
pH	6.7	6.7 -7.1
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Geobacillus stearothermophilus ATCC®7953	<10000	Conforms	Good growth, yellow colonies.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.