

CERTIFICATE OF ANALYSIS

PRODUCT PP1068P90
CHOC II + HB

LOT NUMBER 2975378

EXPIRY DATE 07.07.2020

**DATE OF
MANUFACTURE** 28.04.2020

Physical Characteristics	Results	Specification
Appearance	Chocolate brown gel	Chocolate brown gel
pH	7.13	7.00 - 7.40

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
CO₂, 35°C, 24-48hr		
<i>Haemophilus influenzae</i> ATCC®10211	1-3mm colourless colonies	
<i>Neisseria gonorrhoeae</i> ATCC®43069	1-3mm colourless colonies	

Released By



Nurliyana binti Nazri

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com