

# CERTIFICATE OF ANALYSIS

**PRODUCT** PP1273P90  
SS

**LOT NUMBER** 911712

**EXPIRY DATE** 19.07.2010

**DATE OF MANUFACTURE** Un.no.n

---

**Physical Characteristics**

**Results**

**Specification**

---

Appearance	Pink-red coloured gel	Pink-red coloured gel
pH	7.04	6.80 - 7.20

---

**Sterility:** Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

---

**Microbiological Characteristics**

**Results**

**Specification Quantitative Tests**

**O2, 35°C, 24hr**

*Salmonella typhimurium* ATCC®14028

1-2mm colourless cols with black centres

*Shigella flexneri* ATCC®12022

1-2mm colourless cols

*Escherichia coli* ATCC®25922


Partial inhibition

*Enterococcus faecalis* ATCC®29212

Complete inhibition

---

**Released By**



Aishah Dahlan

---

**MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)**

**(Formerly known as Biomedica Laboratories Sdn Bhd)**

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com