

CERTIFICATE OF ANALYSIS

PRODUCT PP1355P90
XLD

LOT NUMBER 2527858

EXPIRY DATE 02.07.2019

**DATE OF
MANUFACTURE** 07.05.2019

Physical Characteristics	Results	Specification
Appearance	Red, slightly opalescent gel	Red, slightly opalescent gel
pH	7.35	7.20 - 7.60

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
O2, 35°C, 24hr		
<i>Salmonella typhimurium</i> ATCC®14028	1-2mm pink colonies with black centres	
<i>Shigella flexneri</i> ATCC®12022	1-2mm pink colonies	
<i>Escherichia coli</i> ATCC®25922	Complete inhibition	
<i>Enterococcus faecalis</i> ATCC®29212	Complete inhibition	

Released By



Nurliyana binti Nazri

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com